

Declaration of Conformity

In accordance to EN ISO/IEC 17050-1:2010



We, Esco Micro Pte. Ltd.
of 19 Changi South Street 1
Singapore, 486779

declare on our sole responsibility that the product:

Category : Micro Centrifuge
Trademark : Versati™
Model : MCV-88-_, MCR-88-_

in accordance with the following directives:

2014/35/EU : Low Voltage Directive and its amending directives
2014/30/EU : Electromagnetic Compatibility (EMC) Directive and its
amending directives
2011/65/EU : Restriction of Hazardous Substances (RoHS) Directive and
its amending directives

has been designed to comply with the requirement of the following Harmonized
Standard:

EN 61010-1	: Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements
EN 61326-1 Class B	: Electrical equipment for measurement, control, and laboratory use – EMC requirements – Part 1: General requirements
EU 2017/746 Annex III	: In Vitro Diagnostic Medical Devices (IVDR)

Relevant EC directives/standards

EN 61010-2-020	: Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-020: Particular requirements for laboratory centrifuges
-----------------------	---

More information may be obtained from Esco Lifesciences' authorized distributors
located within the European Union. A list of these parties and their contact
information is available upon request.

Place of Issue : Singapore
Date of Issue : December 2022

Indra Setiawan

Indra Setiawan
Plant Manager, Esco Lifesciences