

Declaration of Conformity

In accordance to EN ISO/IEC 17050-1:2010



We, Esco Micro Pte. Ltd.
of 19 Changi South Street 1
Singapore, 486779

declare on our sole responsibility that the product:

Category : Tabletop Centrifuge
Trademark : Versati™
Model : TCV-1500-_, TCR-1500-__

in accordance with the following directives:

2014/35/EU : Low Voltage Directive and its amending directives
2014/30/EU : Electromagnetic Compatibility (EMC) Directive and its amending directives
2011/65/EU : Restriction of Hazardous Substances (RoHS) Directive and its amending directives

has been designed to comply with the requirement of the following Harmonized Standard:

| | |
|------------------------------|---|
| EN 61010-1 | : Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements |
| EN 61326-1 Class B | : Electrical equipment for measurement, control, and laboratory use – EMC requirements – Part 1: General requirements |
| EU 2017/746 Annex III | : In Vitro Diagnostic Medical Devices (IVDR) |

Relevant EC directives/standards

| | |
|-----------------------|---|
| EN 61010-2-020 | : Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-020: Particular requirements for laboratory centrifuges |
|-----------------------|---|

More information may be obtained from Esco Lifesciences' authorized distributors located within the European Union. A list of these parties and their contact information is available upon request.

Place of Issue : Singapore
Date of Issue : December 2022

Indra Setiawan

Indra Setiawan
Plant Manager, Esco Lifesciences